

Certificate number: o10203905i286640-A-1



# Calibration Certificate



Customer  
PHYSICS ASSOCIATES  
PETERS CREEK RD STE A-3 5346  
24019 ROANOKE  
UNITED STATES

Laboratory  
Unfors RaySafe AB  
Uggleåsvägen 29  
SE-427 40 Billdal  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 286640  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found Not performed  
As left 2021-06-01  
Adjustment done Yes  
Tested by Kreshnik Raca

Approved by

Jakob Dahlqvist  
Lab Technician

Certificate date

2021-06-02

## CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Air kerma: 5%

Voltage: 2% or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

